



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
09/449,002	11/24/1999	JOHN P. FALLON	1166/58111	5698

7590 01/22/2007  
RICHARD F. JAWORSKI  
IVAN S. KAVRUKOV  
COOPER & DUNHAM L.L.P.  
1185 AVENUE OF THE AMERICAS 23RD FLOOR  
NEW YORK, NY 11036

EXAMINER
----------

SMITH, RUTH S

ART UNIT	PAPER NUMBER
----------	--------------

3737

SHORTENED STATUTORY PERIOD OF RESPONSE	MAIL DATE	DELIVERY MODE
3 MONTHS	01/22/2007	PAPER

**Please find below and/or attached an Office communication concerning this application or proceeding.**

If NO period for reply is specified above, the maximum statutory period will apply and will expire 6 MONTHS from the mailing date of this communication.

## Office Action Summary

Application No.

09/449,002

Applicant(s)

FALLON ET AL.

Examiner

Ruth S. Smith

Art Unit

3737

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --  
**Period for Reply**

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

### Status

- 1) ☒ Responsive to communication(s) filed on 20 October 2006.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

### Disposition of Claims

- 4) ☒ Claim(s) 23-41 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) ☒ Claim(s) 23-30 is/are allowed.
- 6) ☒ Claim(s) 31-41 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

### Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

### Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some \* c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
  - ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

### Attachment(s)

- ☐ Notice of References Cited (PTO-892)
- ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- ☐ Information Disclosure Statement(s) (PTO/SB/08)  
Paper No(s)/Mail Date \_\_\_\_\_
- ☐ Interview Summary (PTO-413)  
Paper No(s)/Mail Date. \_\_\_\_\_
- ☐ Notice of Informal Patent Application
- ☐ Other: \_\_\_\_\_

***Claim Rejections - 35 USC § 102***

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(e) the invention was described in a patent granted on an application for patent by another filed in the United States before the invention thereof by the applicant for patent, or on an international application by another who has fulfilled the requirements of paragraphs (1), (2), and (4) of section 371(c) of this title before the invention thereof by the applicant for patent.

The changes made to 35 U.S.C. 102(e) by the American Inventors Protection Act of 1999 (AIPA) do not apply to the examination of this application as the application being examined was not (1) filed on or after November 29, 2000, or (2) voluntarily published under 35 U.S.C. 122(b). Therefore, this application is examined under 35 U.S.C. 102(e) prior to the amendment by the AIPA (pre-AIPA 35 U.S.C. 102(e)).

Claims 31-33,37-39 are rejected under 35 U.S.C. 102(e) as being anticipated by Mortimore et al. The claims are directly readable on Mortimore et al in that the quantitative data is considered to be placed in a field other than an image field. It should be noted that applicant states that DICOM is the standard used in the industry to communicate medical files in electronic form. Mortimore et al disclose the use of a DICOM compatible file, see column 3, line 52, column 4, lines 39-41, column 5, line 28.

Claims 31-33,37-39 are rejected under 35 U.S.C. 102(e) as being anticipated by Finger. The claims are directly readable on Finger in that the quantitative data is considered to be placed in a field other than an image field. It should be noted that applicant states that DICOM is the standard used in the industry to communicate medical files in electronic form. Finger discloses the use of a DICOM compatible file, see column 22, lines 47-50.

***Claim Rejections - 35 USC § 103***

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

Art Unit: 3737

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 34-36,40,41 are rejected under 35 U.S.C. 103(a) as being unpatentable over Finger or Mortimore et al. Mortimore et al discloses a method for generating a DICOM compatible file where the quantitative data is considered to be placed in a field other than an image field. It should be noted that applicant states that DICOM is the standard used in the industry to communicate medical files in electronic form. Mortimore et al disclose the use of a DICOM compatible file, see column 3, line 52, column 4, lines 39-41, column 5, line 28. Finger discloses a method for generating a DICOM compatible file where the quantitative data is considered to be placed in a field other than an image field. It should be noted that applicant states that DICOM is the standard used in the industry to communicate medical files in electronic form. Finger discloses the use of a DICOM compatible file, see column 22, lines 47-50. The use of a bone densitometer to provide a bone scan and its related data is old and well known, as disclosed by applicant. It would have been obvious to one skilled in the art to have modified Finger or Mortimore such that the image data provided is provided by a bone densitometer and the quantitative data being related to the bone scan. Such a modification merely involves the substitution of one known type of scan for another. With regard to claims 36,40,41, the use of HTML, XML or Java Script files and the transmission of such across a network (as disclosed by Finger) is well known and the use of such in the communication of medical data would have therefore been obvious to one skilled in the art.

#### ***Allowable Subject Matter***

Claims 23-30 are allowable over the prior art of record.

***Response to Arguments***

Applicant's arguments with respect to claims 31-41 have been considered but are moot in view of the new ground(s) of rejection.

***Conclusion***

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Ruth S. Smith whose telephone number is 571-272-4745. The examiner can normally be reached on M-F 7:30 AM-4:00 PM.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Brian Casler can be reached on 571-272-4956. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3737

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

A handwritten signature in black ink, appearing to read 'Ruth S. Smith', with a stylized, cursive script.

Ruth S. Smith  
Primary Examiner  
Art Unit 3737

RSS